

BROOKFIELD SPIRAL ADAPTER

Operating Instructions

Manual No. **M93-270-E0514**



with offices in: Boston • Chicago • London • Stuttgart • Guangzhou

Distributed by:

Lab Unlimited
CARL STUART GROUP

Tallaght Business Park
Whitestown, Dublin 24,
Ireland
D24 RFK3

Tel: (01) 4523432
Fax: (01) 4523967
E-mail: info@labunlimited.com
Web: www.labunlimited.com

Quatro House, Frimley Road,
Camberley,
United Kingdom
GU16 7ER

Tel: 08452 30 40 30
Fax: 08452 30 50 30
E-mail: info@labunlimited.co.uk
Web: www.labunlimited.co.uk

TABLE OF CONTENTS

I. Introduction	5
II. Components.....	6
III. Specifications for Test Set Up	7
IV. Installation	8
V. Operation	10
V.1 Theory	10
V.2 Test Procedure.....	10
V.3 Data Gathering	10
V.4 Application Examples - Solder Paste	11
V.5 Cleaning the Spiral Adapter	11
Appendix A - Spiral Spindle Factors for Torque Measurements.....	12
Appendix B - Online Help and Additional Resources	13
Appendix C - Warranty Repair and Service.....	14

I. INTRODUCTION

When used with any Standard Brookfield Viscometer/Rheometer, the Spiral Adapter can measure fluid viscosity at various shear rates. Viscosity is a measure of a fluid's resistance to flow. You will find a detailed description of the mathematics of viscosity in the Brookfield publication *"More Solutions to Sticky Problems"*, a copy of which was included with your Viscometer/Rheometer and is also available as a pdf file on the Brookfield website.

The Spiral Adapter has an inner, threaded spindle surrounded by a concentric outer cylinder, as shown in **Figure 1**. This combination causes the sample to be continuously pumped up through the gap between the rotating spindle and the outer cylinder. The material reaches a steady state of flow during which viscosity is measured.




Figure 1

The principle of viscosity measurement is to drive the spiral spindle (which is immersed in the test fluid) through a calibrated spring in the Viscometer/Rheometer. The viscous drag of the fluid against the spindle is measured by the spring deflection.

The measuring range of the Spiral Adapter [in centipoise (cP) or milliPascal seconds (mPa•s)] is determined by the rotational speed of the spindle, the length and effective diameter of the spindle, the length and diameter of the chamber the spindle is rotating in, and the full scale torque of the viscometer's calibrated spring.

II. COMPONENTS

The following components are included in the Spiral Adapter Accessory (part number **SAA**):

<u>Item</u>	<u>Part Number</u>
Two (2) Spindles of identical design	SA-70
One (1) Chamber	SA-1Y
One (1) Assembly Clamp	SA-5Y
One (1) Carrying Case	SA-8
One (1) Cleaning Brush	SA-4
One (1) Operator Manual	M93-270

Please check to be sure that you have received all components, and that there is no damage. If any parts are missing, please notify Brookfield Engineering or your local authorized Brookfield dealer immediately. Any shipping damage must be reported to the carrier.

III. SPECIFICATIONS FOR TEST SET UP

The Spiral Adapter can be used with any Standard Brookfield Viscometer/Rheometer. Selection of the appropriate Viscometer/Rheometer depends upon the viscosity range of the sample fluid over the shear rates at which it will be tested.

There are four basic spring torque series offered by Brookfield as shown in **Table 1**. The higher the torque range, the higher the viscosity measurement range. The corresponding ranges for viscosity measurement when the spiral spindle is rotating at 1.0 RPM are also shown. Lower spindle speeds are possible, but the pumping action within the spiral chamber may not be sufficient to provide meaningful viscosity data.

Brookfield Viscometer/Rheometer Series Designation	Spring Torque		Spiral Adapter Viscosity Range at 1.0 RPM cP (mPa·s)	
	dyne-cm	milli Newton-m	minimum	maximum
LV	673.7	0.0673	1.0×10^4	1.0×10^5
RV	7,187.0	0.7187	1.1×10^5	1.1×10^6
HA	14,374.0	1.4374	2.2×10^5	2.2×10^6
HB	57,496.0	5.7496	9.0×10^5	9.0×10^6

Table 1

Table 2 provides general information on Brookfield Viscometers/Rheometers and the type of data that can be generated when using the Spiral Adapter.

Measurement accuracy for any viscosity value obtained during a test is $\pm 2.0\%$ of the full scale viscosity range (i.e. at 100% torque) for the specific rotational speed in use. See **Appendix A** for the necessary information to determine full scale range.

Model	Dial	DV-E	DV-I Series	DV-II/2T Series	DV-III/3T Series
Speeds	8	18	18	54	2,600
Shear Rate	0.677 sec ⁻¹ per RPM				
Signal Options					
Torque	No	No	Yes	Yes	Yes
Viscosity	No	No	No	Yes	Yes
Temperature	No	No	*	Yes	Yes
Software	No	No	No	Yes	Yes
Temperature Probe	No		*	Yes	

Table 2

Note: At speeds of 1 RPM and lower, additional time may be required to allow for complete deflection of the torque sensor.

* Available as option when ordering instrument.

IV. INSTALLATION

Install the spiral spindle by lifting and holding the pointer shaft of the Viscometer/Rheometer while screwing the spindle onto it. The threads are left-hand so the spindle must be turned to the left. Gently lift the chamber over the spindle up against the pivot cup until it rests over the locating shoulder. While holding the chamber in place, fasten the assembly clamp onto the pivot cup, thereby securing the chamber in place. The thread in this case is right hand.

See **Figure 2** which shows an assembly schematic for the Spiral Adapter.




Figure 2

Turn on the Viscometer/Rheometer and rotate the spindle at various speeds. Observe the spiral spindle to make sure there is no interference with steady and even rotation. If the spindle does not run “true”, then inspect the spindle and Viscometer/Rheometer shaft for straightness. Rotate the Viscometer/Rheometer shaft without the spindle and confirm that the rotation is steady and even.

Notify Brookfield or your local authorized Brookfield dealer if you experience any difficulty in performing the above procedure.

The spiral spindle code number is **70**. If you have a Brookfield digital viscometer or rheometer, which does not contain this number, you must do the following:

- For DV-I and Dial Viscometers, use the torque reading and manually convert to a viscosity value using the table in **Appendix A**.
- For DV-II Viscometers, use the spindle **99** entry. Enter the following values:
SMC (Spindle Multiplier Constant) = 105
SRC (Shear Rate Constant) = .677

Note: Customers who use the Brookfield Spiral Adapter may need to have their Brookfield Viscometer/Rheometer serviced before using this accessory device. All Brookfield digital viscometers/rheometers with serial number lower than 24250 may not have proper alignment between bearings and pivot cup. The same applies to all Brookfield dial viscometers.

This does not present any problem for normal Brookfield Viscometer/Rheometer use. However, it may result in erroneous data when using the Spiral Adapter. For this reason, please return your Brookfield Viscometer/Rheometer to Brookfield for servicing before using it with the Spiral Adapter.

V. OPERATION

V.1 Theory

The Spiral Adapter has an inner, threaded spindle surrounded by a concentric outer cylinder, as shown previously in **Figure 1**. This combination causes the sample to be continuously pumped up through the gap between the rotating spindle and the outer cylinder. The material reaches a steady state of flow during which viscosity is measured.

At its various rotating speeds, the Brookfield Viscometer/Rheometer with the Spiral Adapter provides viscosity data at different shear rates. The resulting rheogram from an up/down speed ramp supplies important information on the pseudoplastic and thixotropic behavior of the test material which may correlate with actual use in a process. In addition, the measurement method may be less sensitive to sample handling and minor material variations than other viscosity measuring methods.

V.2 Test Procedure

Ensure that the test material has been prepared properly in accordance with your prescribed procedure. Lower the Spiral Adapter into the test fluid as shown in **Figure 1**.

Start the Viscometer/Rheometer and allow it to run at high speed for quick flooding of the chamber. This speed will vary with the nature of the test fluid, but will typically be between 20 and 60 RPM for 2 to 10 minutes. The speed and time of the flood cycle may be an important part of the test procedure. This is because the fluid in the chamber will experience a shear history before the start of the test.

When the chamber is flooded, you will observe test fluid covering the top of the spindle. Stop the Viscometer/Rheometer and allow a settling period per your defined procedure. Run the viscosity test and record the data. An example of a test procedure is as follows:

<u>Speed</u>	<u>Duration</u>
50 RPM	6 min.
0 RPM	2 min.
5 RPM	1 min.
10 RPM	1 min.
15 RPM	1 min.
20 RPM	1 min.
30 RPM	1 min.

V.3 Data Gathering

Brookfield offers applications software for comprehensive data gathering with any IBM compatible PC.

Brookfield software will capture and display the following data in tabular and/or graphical format when used with the appropriate Brookfield Viscometer.

- Viscosity in centipoise (cP) or milliPascal seconds (mPa•s)
- % scale (Brookfield units)
- RPM

- Shear rate (sec^{-1})
- Shear stress (dynes/cm^2 or N/m^2)
- Temperature ($^{\circ}\text{F}$ or $^{\circ}\text{C}$)

The data from a Brookfield dial reading, DV-E or DV-I+ (or DV-I) Viscometer can be manually entered into one of our software applications.

Contact Brookfield Engineering or your local authorized dealer for more details on the appropriate software to use with your viscometer/rheometer.

V.4 Application Example - Solder Paste

Applying solder paste is a printing operation that shears the materials as the paste “rolls” on the stencil surface. Solder paste typically shows pseudoplastic behavior (exponential decrease in viscosity with increasing shear rate) as shown in **Figure 4**.




Figure 4




Figure 5

A plot of viscosity vs. RPM on a log/log scale approximates a straight line as shown in **Figure 5**. The slope of the line provides information about the shear thinning of the paste. The absolute value of the slope is called the “shear sensitivity factor”. Experimental correlations of these slope values to actual printing properties allow the user to establish acceptable ranges of the shear sensitivity factor.

The plots shown in **Figures 4** and **5** can be generated easily when using Brookfield application software. There is a math model (IPC Paste Analysis) included in the software, which converts viscosity vs. RPM data (**Figure 4**) into a log/log plot (**Figure 5**) and automatically computes the “shear sensitivity factor.”

V.5 Cleaning the Spiral Adapter

Support the chamber and remove the clamp. Carefully lower the chamber off the spindle. Soak in appropriate solvent. Remove spindle by lifting the lower shaft and holding it firm. Turn the spindle nut to the right (counter-clockwise) to remove. Clean by wiping and soaking in appropriate solvent. A soft brush is provided to aid in cleaning the threads. This brush should also be cleaned in solvent.

APPENDIX A - Spiral Spindle Factors for Torque (% Scale) Measurements

Table A-1 applies to Brookfield Viscometer models with series designation LV, RV, HA, and HB. The procedure for converting a torque value (% scale) into a viscosity value (cP) is accomplished by multiplying the appropriate factor shown in **Table A-1** by the torque reading on your Brookfield Viscometer.

Spindle Speeds (RPM)	Spiral Spindle Factors			
	LV	RV	HA	HB
0.3	3,280.0	35,000	70,000	280,000
0.5	1,968.0	21,000	42,000	168,000
0.6	1,640.0	17,500	35,000	140,000
1.0	984.0	10,500	21,000	84,000
1.5	656.0	7,000	14,000	56,000
2.0	492.0	5,250	10,500	42,000
2.5	394.0	4,200	8,400	33,600
3.0	328.0	3,500	7,000	28,000
4.0	246.0	2,625	5,250	21,000
5.0	197.0	2,100	4,200	16,800
6.0	164.0	1,750	3,500	14,000
10.0	98.4	1,050	2,100	8,400
12.0	82.0	875	1,750	7,000
20.0	49.2	525	1,050	4,200
30.0	32.8	350	700	2,800
50.0	19.7	210	420	1,680
60.0	16.4	175	350	1,400
100.0	9.8	105	210	840

Table A-1

The full scale viscosity measurement range (in cP or mPa*s) for any given speed is obtained by multiplying the corresponding spiral spindle factor in the above table by 100.

APPENDIX B - Online Help and Additional Resources

www.brookfieldengineering.com**

The Brookfield website is a good resource for additional and self-help whenever you need it. Our website offers a selection of “how-to” videos, application notes, conversion tables, instructional manuals, material safety data sheets, calibration templates and other technical resources.

<http://www.youtube.com/user/BrookfieldEng>

Brookfield has its own YouTube channel. Videos posted to our website can be found here as well as other “home-made” videos made by our own technical sales group.

Viscosityjournal.com

Brookfield is involved with a satellite website that should be your first stop in viscosity research. This site serves as a library of interviews with experts in the viscosity field as well as Brookfield technical articles and conversion charts. Registration is required, so that you can be notified of upcoming interviews and events, however, this information will not be shared with other vendors, institutions, etc..

Article Reprints

Available in Print Only

- Brookfield has an extensive library of published articles relating to viscosity, texture and powder testing. Due to copyright restrictions, these articles cannot be emailed. Please request your hardcopy of articles by calling our customer service department directly or by emailing: marketing@brookfieldengineering.com

Available Online

- Brookfield has a growing number of published articles that can be downloaded directly from the Brookfield website. These articles can be found on our main site by following this path: <http://www.brookfieldengineering.com/support/documentation/article-reprints>

More Solutions to Sticky Problems

Learn more about viscosity and rheology with our most popular publication. This informative booklet will provide you with measurement techniques, advice and much more. It's a must-have for any Brookfield Viscometer or Rheometer operator. More Solutions is available in print and also as a downloadable pdf on the Brookfield website by following this path: <http://www.brookfieldengineering.com/support/documentation>

Training/Courses

Whether it is instrument-specific courses, training to help you better prepare for auditing concerns, or just a better understanding of your methods, who better to learn from than the worldwide leaders of viscosity measuring equipment? Visit our Services section on our website to learn more about training.

** Downloads will require you to register your name, company and email address. We respect your privacy and will not share this information outside of Brookfield.

APPENDIX C - Warranty Repair and Service

The Brookfield Spiral Adapter is guaranteed for one year from date of purchase against defects in materials and workmanship. The Spiral Adapter must be returned to **Brookfield Engineering Laboratories, Inc.** or to the Brookfield dealer from whom it was purchased for a warranty evaluation. Transportation is at the purchaser's expense. The Spiral Adapter should be shipped together with all components originally provided with the system. If returning to Brookfield, please contact us for a return authorization number prior to shipping.

For a copy of the Repair Return Form, go to the Brookfield website,
www.brookfieldengineering.com

For repair or service in the **United States** return to:

Brookfield Engineering Laboratories, Inc.
11 Commerce Boulevard
Middleboro, MA 02346 U.S.A.

Telephone: (508) 946-6200 FAX: (508) 923-5009
www.brookfieldengineering.com

For repair or service outside the United States consult Brookfield Engineering Laboratories, Inc. or the dealer from whom you purchased the instrument.

For repair or service in the **United Kingdom** return to:

Brookfield Viscometers Limited
1 Whitehall Estate
Flex Meadow, Pinnacles West
Harlow, Essex CM19 5TJ, United Kingdom

Telephone: (44) 27/945 1774 FAX: (44) 27/945 1775
www.brookfield.co.uk

For repair or service in **Germany** return to:

Brookfield Engineering Laboratories Vertriebs GmbH
Hauptstrasse 18
D-73547 Lorch, Germany

Telephone: (49) 7172/927100 FAX: (49) 7172/927105
www.brookfield-gmbh.de

For repair or service in **China** return to:

Guangzhou Brookfield Viscometers and Texture Instruments Service Company Ltd.
Room C1, 5/F, Tianxing Building East Tower, No. 21, Zhongshan Yi Road, Yuexiu
District
Guangzhou, 510600, P. R. China

Telephone: (86) 20/3760-0548 FAX: (86) 20/3760-0548
www.brookfield.com.cn

On-site service at your facility is also available from Brookfield. Please contact our Service Department in the United States, United Kingdom, Germany or China for details.